



Temperature control equipment for viscosity measurements for viscometers "ST-2020" and "STS-2011"

The influence of temperature while measuring viscosity is considerable, lower temperatures increase viscosity. Therefore it is indispensable to control temperature when precise viscosity measurements are required. The control of temperature by using a thermostatically controlled bath is the most efficient, because of the re-circulation of liquid produces a rapid and stable temperature.

We recommend our range of immersion thermostats for this application.

TEMPERATURE CONTROL AND CONFIGURATION

FOR VISCOSITY MEASUREMENTS:

FRIGITERM-TFT-10 Part No. **3000546**, **FRIGTERM-TFT-30** Part No. **3000547**

Suitable for below ambient working environments (see page 101 for specifications).

For temperatures above ambient see the **DIGITERM-S-150** Part No. **3000543**, or **DIGITEM-TFT-200** Part No. **3000544** complete with 12 litre tank Part No **6000391**, (see pages 98 and 100 for specifications).

An adapter kit for the thermostat bath is required:

Part No. **1001625** for STS-2011

Part No. **1001626** for ST-2020

Adapter for thermostat bath comprising of an extension spindle and 4 leg adjusters for the bath. (Kit can be fitted by the user.)

This kit adjusts the bath measuring height to enable samples to be placed inside.

Part No. **1001627**. Support for 2 500 ml beakers, to be placed inside the bath (only for bucket **6000391**).

Part No. **1001628** Support base for the **DIGITERM** immersion thermostats. This accessory is recommended to close the bath and maintain a constant temperature.

CONSTANT TEMPERATURE CONTROL ACCESSORY FOR SMALL SAMPLE VOLUME ADAPTERS. USE WITH THE RE-CIRCULATING JACKET ACCESSORIES.

PART NO. **1000996**


To work below ambient temperatures, we recommend the use of the **FRIGITERM-TFT-10** Part No. **3000546**, or **FRIGTERM-TFT-30** Part No. **3000547** (see page 101 for specifications.)

For temperatures above ambient see the **DIGITERM-S-150** Part No. **3000543**, or **DIGITEM-TFT-200** Part No. **3000544** complete with 12 litre tank Part No. **6000391**, (see pages 98 and 100 for specifications).

It is necessary to configure the immersion thermostats for "external re-circulation." The Bath adapter kit is not required.



Frigiterm thermostat bath with adapter kit
Part No. 1001625, fitted with a rotary viscometer.



Digiterm thermostat bath with support base Part No. 1001628, 12 litre bath Part No. 6000391 and adapter kit Part No. 1001625 or 1001626 To be used with a rotary viscometer and beaker support Part No. 1001627.



Immersion thermostat Digiterm controlling temperature of viscometer.



COMECTA **Rotary Viscometer "NDJ-1"**

APPLICATIONS

The instrument operates by rotating a disk or cylinder, (spindle), that is submerged in the liquid or fluid to be analysed. A pre-selected speed is set, the unit measures the absolute resistance from viscosity of the fluid being analysed. Suitable for samples such as: foods, cosmetics, fats and oils, pharmaceutical products, paints and plastics, etc.

FEATURES

The asynchronous motor is connected to a graduated disk with 4 different speeds that propel the spindle via a spiral and die.

Supplied complete with a set of 4 spindles in a box, numbered 1 to 4 with spindle stand.

The viscometer includes a level and adjustable screw feet support base and protective case.

MODEL	Part No.	Measuring range	Tolerance	Spindle Speed r.p.m. for 1 to 4	Power W	Weight Kg
NDJ-1	5120230	10-100.000 mPaS	±5% Liquid Newtons	6 - 12 - 30 - 60	15	6